



| Biological Indicator Steam |                                       |
|----------------------------|---------------------------------------|
| Model                      | UGBI0501                              |
| Packing                    | 50pcs/Box                             |
| Organism                   | G-stearothermophilus ATCC7953         |
| Population(mean/strip)     | ≥1.0x10 <sup>5</sup> c.f.u.           |
| D-value                    | ≥1.5 min(121°C)                       |
| Incubation Time            | 24~48H                                |
| Frequency of Use           | Test 1 time every sterilization cycle |